

Calibration

For all Thyracont vacuum gauges



**TESTED &
APPROVED**

Calibrations

- Each sensor is provided with individual temperature compensation that is applied using a specially developed 4 point calibration method. These calibrations are traceable to national DAkkS standards for maximum measuring precision. The gauges are generally delivered without a calibration certificate.
- A works calibration by Thyracont with a 4-page documentation can be ordered with new gauges as a service (ADJCERT, ADJCERTHV). We can also offer a DAkkS calibration by a third party institute (DKDCERT, DKDCERTHV).
- Depending on the application, we suggest to have your gauges calibrated in annual intervals. A service contract with the duration of 5 years includes besides the yearly examination and works calibration also a reminder service (WVADJCERT).

Article numbers

- ADJCERT Works calibration, 4 ref. points per pressure decade, 1000 - 1e-4 mbar
 ADJCERTHV Works calibration, 4 ref. per pressure decade, 1000 - 1e-5 mbar
- WVADJCERT as ADJCERT with contract for 5 years, annual calibration
 WVADJCERTHV wie ADJCERTHV with contract for 5 years, annual calibration
- DKDCERT DAkkS calibration, 1000 - 1e-4 mbar
 DKDCERTHV DAkkS calibration, 1000 - 1e-6 mbar